

Validation Report #029635

Validation Date: 03/19/14

Summary

Antigen	C-Telopeptide of Type I Collagen
Catalog number	ABIN367349
Supplier	Cusabio
Supplier catalog number	csb-e10363h
Lot number	2014-2AX-2014-08AX
Method validated	Enzyme-linked immunosorbent assay
Laboratory	Affina Biotechnologies, Inc
Validation number	029635
Positive Control	Human serum
Negative Control	Chicken plasma
Notes	Signal was detected in positive control sample and not in negative control sample.



Full Methods

Primary Antibody

- Antibody: C-Telopeptide of Type I Collagen (ICTP)
- Catalog number: ABIN367349
- Supplier: Cusabio
- Supplier catalog number: csb-e10363h
- Lot number: 2014-2AX-2014-08AX

Controls

- Positive control: Human individual post-menopausal female serum (Biochemed, 750-NS-FI-POM) (positive)
- Negative control: Chicken plasma (Sigma-Aldrich, p3266) – reconstituted at 1 mg/mL
- Standard curve: 0, 25, 75, 175, 400, 800 ng/mL ICTP provided in the ELISA kit
- Spike control: 800 ng/mL standard premixed with chicken plasma in a 1:1 ratio

Protocol

- 50 μ L of standard and samples were added 48-well strip plates provided in the kit with 50 μ L of HRP conjugate. All samples and standards were assayed in duplicate.
- The microplate was covered and incubated at 37°C for 1 h.
- Plate contents were discarded.
- Content of the wells was discarded and wells were washed 3 times with 200 μ L of distilled water with a 1 min soak for each wash.
- 50 μ L of Substrate A and Substrate B each was added to each well. The plate was covered and incubated at 37°C for 15 min.
- 50 μ L of the Stop Solution was added per well.
- The entire sample was transferred into a 96-well plate (Nunc, Maxisorp)
- The optical density (OD value) of each well was read immediately using a microplate reader set to 450 nm.
- The duplicate readings for each sample were averaged and the average zero standard optical density subtracted. The corrected average-value was tabulated as Average Absorbance. A standard curve was generated by plotting the mean OD value for each standard on the x-axis against the concentration on the Y-axis using CurveExpert 1.4. A linear equation was used for the best fit through the points on the graph.
- The CurveExpert Analyze feature was used to calculate human ICTP concentrations of the samples based on their Average Absorbance values.

Experimental Notes

- Transfers of final samples for OD reading were necessitated by inability to read a 48-well plate in the standard plate carriage. This caused high variance between duplicates.

Figures




Figure 1: Graph of corrected-average absorbance (OD 450 nm) readings plotted for standard curve samples.

Equation: $OD\ 450 = 5.00 + 1.804 * \text{concentration}$

Type	Sample ng/ml	Reading-1	Reading-2	Avg Reading	Avg Absorbance	SD	Calculated Conc
Standard Curve	800	1.710	1.224	1.467	1.428	0.028	763.8
	400	1.014	0.839	0.927	0.888	0.028	464.6
	175	0.504	0.415	0.460	0.421	0.028	205.8
	75	0.149	0.199	0.174	0.135	0.028	47.3
	25	0.139	0.116	0.128	0.089	0.028	21.9
	blank	0.040	0.038	0.039			
Spike Control	400	1.157	1.059	1.108	1.069	0.070	565.0
Positive Control	Human serum (1/2 diluted)	0.623	0.584	0.604	0.5645	0.028	285.9
Negative control	Chicken Plasma	0.040	0.039	0.040	0.001	0.001	-27.0

Table 1: Table of absorbance readings (OD 450 nm) for standard curve, spike controls and unknown control samples. Value for Average Reading is derived from the average of two readings (OD 450 nm). The Average Reading for blank sample (no conjugate added) was subtracted from all Average Readings to yield Average Absorbance values for Standards, spike controls and control samples. Standard deviation is included for all samples. The concentration of samples was calculated using the Analyze feature of the CurveExpert 1.4 software for a linear equation fit.